PI 0312, Rev. B DCN: 3746

Mohs Mouse/Rabbit PolyDetector HRP Green Detection System







www.biosb.com

Intended Use

For Mohs In Vitro Diagnostic Use.

This detection system is intended for the fast immunohistochemical detection of antibodies used for Mohs surgery on frozen or FFPE tissue sections. Interpretation of results should be performed by a qualified medical professional.

Summary and Explanation

The Mohs Mouse/Rabbit PolyDetector HRP Green Detection System is a, 1-step Fab micropolymeric detection system that allows for the demonstration of antigens in cryostat sections, formalin-fixed paraffin-embedded tissue, blood smears, cytosmears, and cell preparations. The Mohs PolyDetector kit has been developed using a biopolymer conjugated to monomeric Fab' immunoglobulin fractions targeting the Fc region of Mouse and Rabbit antibodies. Additionally, the biopolymer is labeled with high quality HRP for maximum sensitivity. This ensures excellent cellular penetration which generates consistent, reproducible sensitive and specific immunostainings for all types of nuclear, cytoplasmic and membranous antigens, in different types of frozen, FFPE tissues and cell preparations.

The increased sensitivity of the Mohs Mouse/Rabbit PolyDetector HRP Green Detection System allows for rapid staining procedures without compromising stain quality. The Mohs Mouse/Rabbit PolyDetector HRP Green Detection System is suitable for use with mouse IgG and IgM and rabbit primary antibodies, both monoclonal and polyclonal. The Mohs Mouse/Rabbit PolyDetector HRP Green Detection System kits are optimized for use with Bio SB TintoFast primary antibodies; however, they are universal kits and therefore should work equally well with antibodies from different vendors, as long as they are properly optimized.

Precautions

- 1 For professional users only. Results should be interpreted by a qualified medical professional.
- 2. Ensure proper handling procedures are used with reagent. Minimize microbial contamination of reagents.
- 3. Always wear proper personal protective equipment such as laboratory coat, goggles and gloves when handling reagents.
- 4. Dispose of unused solution according to local and federal regulations.
- $5.\,Do\,not\,ingest\,reagent.\,If\,reagent\,is\,ingested, seek\,medical\,advice\,immediately.$
- 6. Avoid contact with eyes. If contact occurs, flush with large quantities of water.
- 7. Follow safety precautions for the heating device used for epitope retrieval (TintoRetriever Pressure Cooker (BSB 7008), TintoRetriever PT Module (BSB 7030 or 7033) or similar).
- 8. For additional safety information refer to Safety Data Sheet for this product.
- 9. For complete recommendations for handling biological specimens please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (1).

Presentation

The Mohs Mouse/Rabbit PolyDetector HRP Green Detection System contains a Peroxidase Blocker solution, Anti-Mouse/Rabbit Horseradish Peroxidase micropolymer solution, an HRP Green Buffer Substrate, and an HRP Green Chromogen solution. All components are buffered with stabilizers and a non-azide anti-microbial agent.

Catalog No.	Volume/Qty		
BSB 0310S	5 mL Each		
BSB 0310	15mL Each		
BSB 0311	50 mL Each		
BSB 0312	100 mL Each		

Storage

Store at 2-8°C

Stability

The Mohs Mouse/Rabbit PolyDetector HRP Green Detection System is stable up to the expiration date listed on the product label. Do not use this product after the expiration date listed on the product label. Temperature fluctuations should be avoided. Store appropriately when not in use and avoid prolonged exposure to room temperature conditions.

Preparation of Working Solution

The PolyDetector Peroxidase Blocker and Anti-Mouse/Rabbit Horseradish Peroxidase Label are ready-to-use working solutions and require no further preparation. The HRP Green Chromogen is concentrated and needs to be diluted with the buffer solution (1 drop chromogen/1 ml buffer). Prepare the working Substrate Chromogen Solution up to 6 hours before use.

Specimen Preparation

Paraffin sections: This product can be used on formalin-fixed paraffin-embedded (FFPE) tissue sections. Ensure tissue undergoes appropriate fixation for best results. Pre-treatment of tissues with heat-induced epitope retrieval (HIER) is recommended using Bio SB ImmunoDNA Retriever with Citrate (BSB 0020 - BSB 0023), ImmunoDNA Retriever with EDTA (BSB 0030-BSB 0033). Additionally, TintoDeparaffinator Citrate or EDTA (BSB 0175 - BSB 0178) can be used to deparaffinize, retrieve and hydrate FFPE Tissues. Tissue should remain hydrated using Bio SB Immuno/DNA Washer solutions (BSB 0029 & BSB 0042).

Frozen sections and cell preparations: This product can be used for labeling acetone-fixed frozen sections and acetone-fixed cell preparations.

Mohs IHC Procedure

Specimen Preparation of Mohs Frozen Tissues

- 1. Embed the specimen in OCT inside a cryostat.
- 2. Cut sections at 4-5 μ m and mount on a positively charged glass slide such as the Bio SB Hydrophilic Plus Slides (BSB 7028).
- 4. Air dry the slide at room temperature for 2 minutes and then incubate the slide at 60 °C for 3 minutes in an incubator or dry bath.
- 5. Fix in 100% acetone or 10% NBF for 2 minutes at room temperature.
- 6. Rinse with distilled water and air dry the slides for another 2 minutes at room temperature.

Tissue Pretreatment Procedure for Mohs Frozen Tissues

1. Subject tissues to epitope retrieval using Bio SB ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023), ImmunoDNA Retriever with EDTA (BSB 0030-BSB 0033) or Mohs ImmunoDigestor (BSB 0324- 0326). It is recommended that user choose either a 5 minutes heat-induced epitope retrieval (HIER) method using a pressure cooker (BSB TintoRetriever Pressure Cooker) at 110 °C, or a 1-minute proteolytic induced epitope retrieval (PIER) method (for Cytokeratins using 10% NBF fixed tissues). Morphology is better with 10% NBF fixed tissues, especially when using proteolytic induced epitope retrieval method, however, antigen maybe easier to detect on the 100% acetone fixed tissue.

2. Any of three HIER methods may be used:

a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA, and place on trivet or staining dish support in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high (110° C). Incubate for 5 minutes. Open and immediately transfer slides to room temperature. Cool off for 5 -10 min

b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 10-15 minutes.

c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a steamer, cover and steam for 10-15 minutes.

3. For **PIER**, apply ImmunoDNA Digestor (BSB 0108-0112) at room temperature for 1 min then wash.

IHC Detection Procedure

- 1.. After HIER or PIER, transfer slides to ImmunoDNA washer and let stand for 1-2 minutes.
- 2.. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 3.. Wash slides with ImmunoDNA washer or DI water.
- 4. Continue Mohs IHC detection protocol. Wash slides between each step with ImmunoDNA washer solution.

Specimen Preparation for FFPE Tissues

- 1. Cut and mount 3-5-micron formalin-fixed paraffin-embedded tissues on positively charged slides such as Bio SB Hydrophilic Plus Slides (BSB 7028).
- 2. Air dry for 2 hours at 58° C.

Tissue Pretreatment Procedure for FFPE Tissues

- 1. Deparaffinize and rehydrate tissues.
- 2. Subject tissues to heat induced epitope retrieval (HIER) using a suitable retrieval solution such as ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023) or EDTA (BSB 0030-BSB 0033).
- 3. Any of three heating methods may be used:

a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA, and place on trivet in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high. Incubate for 10-15 minutes. Open and immediately transfer slides to room temperature.


b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 20-30 minutes.

c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a steamer, cover and steam for 20-30 minutes.

- 4. After heat treatment, transfer slides in ImmunoDNA Retriever with Citrate or EDTA to room temperature and let stand for 15 minutes.
- 5. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 6. Wash slides with ImmunoDNA washer or DI water.
- 7. Continue IHC staining protocol. Wash slides between each step with ImmunoDNA washer solution.



Abbreviated Immunohistochemical Protocol

Abbi criatca illilianonistotiiciintai i rototoi									
Step	Mohs PolyDetector HRP	Mohs PolyDetector HRP Green 10 min Protocol 0.5 min.							
	Green								
	5 min Protocol								
Peroxidase Blocker	0.5 min.								
Primary Antibody	2 min	4 min.							
1st Step Detection	1 min	3 min.							
Substrate-Chromogen	1 min	2 min.							
Counterstain / Coverslip	0.5 min	0.5 min.							



Mounting Protocols

Mount with aqueous mounting such as AquaMounter (BSB-0090- BSB 0093) or permanent mounting media such as XyGreen PermaMounter (BSB 0169-0174) or organic solvent-based resin such as PermaMounter (BSB 0094-0097).

Product Limitations

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a qualified medical professional.

References

1. U.S. Department of Health and Human Services: Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. Supplement / Vol. 61, January 6, 2012.

Symbol Key / Légende des symboles/Erläuterung der Symbole

				Storage Temperature		Manufacturer		Catalog Number
	V ^{ave}		Limites de température	444	Fabricant	REF	Référence du catalogue	
			2.0	Zulässiger Temperaturbereich		Hersteller		Bestellnummer
	IVD	In Vitro Diagnostic Medical Device		Read Instructions for Use		Expiration Date		Lot Number
	שעו	Dispositif médical de diagnostic in vitro	Consulter les instructions d'utilisation	23	Utiliser jusque	LOT	Code du lot	
		In-Vitro-Diagnostikum		Gebrauchsanweisung beachten	_	Verwendbar bis		Chargenbezeichnung

