

**COMPONENTS****CAS#**

Deionized Water	7732-18-5
Formic Acid	64-18-8
Formaldehyde	50-00-0

EMERGENCY CONTACT: Chemtrec 1-800-424-9300*For In Vitro Diagnostic Use Only*

Fixative Decalcifier **Fix** CalFor[®]

DIRECTIONS FOR USE:

Calfor is formulated to fix and decalcify simultaneously. Fresh tissue can be placed directly into Calfor; Needle biopsies may be completely fixed and decalcified in one hour or less. Small bone specimens and calcified arteries may be completely fixed and decalcified in 2-4 hours. An excellent decalcifier for large dense specimens for fixation and to begin decalcification. Calfor is a proven formulation for ISH and IHC.

DANGER: Harmful if swallowed. Harmful if inhaled. Causes severe skin burns and eye damage. May cause an allergic skin reaction. Obtain special instructions before use. Do not handle until all safety precautions have been read and understood. Use personal protective equipment as required. Wash face, hands and any exposed skin thoroughly after handling. Do not eat, drink or smoke when using this product. Use only outdoors or in a well-ventilated area. Do not breathe dust/fume/gas/mist/vapors/spray. Contaminated work clothing should not be allowed out of the workplace. Wear protective gloves.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or doctor/physician. **IF ON SKIN (OR HAIR):** Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. Wash contaminated clothing before reuse. If skin irritation or rash occurs: Get medical advice/attention. **IF INHALED:** Remove victim to fresh air and keep at rest in a position comfortable for breathing. Call a POISON CENTER or doctor/physician if you feel unwell. Immediately call a POISON CENTER or doctor/physician. **IF SWALLOWED:** Call a POISON CENTER or doctor/physician if you feel unwell. Rinse mouth. Do NOT induce vomiting.

IN CASE OF FIRE: Use CO₂, dry chemical, or foam for extinction.

www.cancerdiagnostics.com

Cancer Diagnostics, Inc. 4300 Emperor Blvd., 400 Durham, NC 27703 (877) 846 5393 Fax: (877) 817 1716

59°F (15°C)  86°F (30°C)

REF CFQ025 1 Gal (3.8L) CUBE

Store locked up at room temperature
Dispose of contents/container to an approved waste disposal plant.



MX1394 CalFor Fixative CUBE AW-26034 Rev.C 2020-07-16

LOT



MADE IN USA

**COMPONENTS****CAS#**

Deionized Water 7732-18-5

Formic Acid 64-18-8

Formaldehyde 50-00-0

EMERGENCY CONTACT: Chemtrec 1-800-424-9300*For Invitro Diagnostic Use Only*

Fixative Decalcifier **Fix** CalFor[®]

DIRECTIONS FOR USE:

Calfor is formulated to fix and decalcify simultaneously. Fresh tissue can be placed directly into Calfor; Needle biopsies may be completely fixed and decalcified in one hour or less. Small bone specimens and calcified arteries may be completely fixed and decalcified in 2-4 hours. An excellent decalcifier for large dense specimens for fixation and to begin decalcification. Calfor is a proven formulation for ISH and IHC.

DANGER: Harmful if swallowed. Harmful if inhaled. Causes severe skin burns and eye damage. May cause an allergic skin reaction. Obtain special instructions before use. Do not handle until all safety precautions have been read and understood. Use personal protective equipment as required. Wash face, hands and any exposed skin thoroughly after handling. Do not eat, drink or smoke when using this product. Use only outdoors or in a well-ventilated area. Do not breathe dust/fume/gas/mist/vapors/spray. Contaminated work clothing should not be allowed out of the workplace. Wear protective gloves.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or doctor/physician. **IF ON SKIN (OR HAIR):** Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. Wash contaminated clothing before reuse. If skin irritation or rash occurs: Get medical advice/attention. **IF INHALED:** Remove victim to fresh air and keep at rest in a position comfortable for breathing. Call a POISON CENTER or doctor/physician if you feel unwell. Immediately call a POISON CENTER or doctor/physician. **IF SWALLOWED:** Call a POISON CENTER or doctor/physician if you feel unwell. Rinse mouth. Do NOT induce vomiting.

IN CASE OF FIRE: Use CO₂, dry chemical, or foam for extinction.

www.cancerdiagnostics.com

Cancer Diagnostics, Inc. 4300 Emperor Blvd., 400 Durham, NC 27703 (877) 846 5393 Fax: (877) 817 1716



REF CFQ038 1 Gal (3.8L)

Store locked up at room temperature
Dispose of contents/container to an approved waste disposal plant.

MX1395 CalFor Fixative 1 Gal AW-26034 Rev.C 2020-07-16

LOT



MADE IN USA

**COMPONENTS****CAS#**

Deionized Water 7732-18-5

Formic Acid 64-18-8

Formaldehyde 50-00-0

EMERGENCY CONTACT: Chemtrec 1-800-424-9300*For In Vitro Diagnostic Use Only*

Fixative Decalcifier **Fix** CalFor[®]

DIRECTIONS FOR USE:

Calfor is formulated to fix and decalcify simultaneously. Fresh tissue can be placed directly into Calfor; Needle biopsies may be completely fixed and decalcified in one hour or less. Small bone specimens and calcified arteries may be completely fixed and decalcified in 2-4 hours. An excellent decalcifier for large dense specimens for fixation and to begin decalcification. Calfor is a proven formulation for ISH and IHC.

DANGER: Harmful if swallowed. Harmful if inhaled. Causes severe skin burns and eye damage. May cause an allergic skin reaction. Obtain special instructions before use. Do not handle until all safety precautions have been read and understood. Use personal protective equipment as required. Wash face, hands and any exposed skin thoroughly after handling. Do not eat, drink or smoke when using this product. Use only outdoors or in a well-ventilated area. Do not breathe dust/fume/gas/mist/vapors/spray. Contaminated work clothing should not be allowed out of the workplace. Wear protective gloves.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or doctor/physician. **IF ON SKIN (OR HAIR):** Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. Wash contaminated clothing before reuse. If skin irritation or rash occurs: Get medical advice/attention. **IF INHALED:** Remove victim to fresh air and keep at rest in a position comfortable for breathing. Call a POISON CENTER or doctor/physician if you feel unwell. Immediately call a POISON CENTER or doctor/physician. **IF SWALLOWED:** Call a POISON CENTER or doctor/physician if you feel unwell. Rinse mouth. Do NOT induce vomiting.

IN CASE OF FIRE: Use CO₂, dry chemical, or foam for extinction.

www.cancerdiagnostics.com

Cancer Diagnostics, Inc. 4300 Emperor Blvd., 400 Durham, NC 27703 (877) 846 5393 Fax: (877) 817 1716



REF CFQ438 1 Gal (3.8L) CS/4

Store locked up at room temperature
Dispose of contents/container to an approved waste disposal plant.

MX1396 CalFor Fixative Gallons, 4 AW-26034 Rev.C 2020-07-16

LOT



MADE IN USA

**COMPONENTS****CAS#**

Deionized Water	7732-18-5
Formic Acid	64-18-8
Formaldehyde	50-00-0

EMERGENCY CONTACT: Chemtrec 1-800-424-9300*For In Vitro Diagnostic Use Only*

Fixative Decalcifier **Fix** CalFor[®]

DIRECTIONS FOR USE:

Calfor is formulated to fix and decalcify simultaneously. Fresh tissue can be placed directly into Calfor; Needle biopsies may be completely fixed and decalcified in one hour or less. Small bone specimens and calcified arteries may be completely fixed and decalcified in 2-4 hours. An excellent decalcifier for large dense specimens for fixation and to begin decalcification. Calfor is a proven formulation for ISH and IHC.

DANGER: Harmful if swallowed. Harmful if inhaled. Causes severe skin burns and eye damage. May cause an allergic skin reaction. Obtain special instructions before use. Do not handle until all safety precautions have been read and understood. Use personal protective equipment as required. Wash face, hands and any exposed skin thoroughly after handling. Do not eat, drink or smoke when using this product. Use only outdoors or in a well-ventilated area. Do not breathe dust/fume/gas/mist/vapors/spray. Contaminated work clothing should not be allowed out of the workplace. Wear protective gloves.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or doctor/physician. **IF ON SKIN (OR HAIR):** Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. Wash contaminated clothing before reuse. If skin irritation or rash occurs: Get medical advice/attention. **IF INHALED:** Remove victim to fresh air and keep at rest in a position comfortable for breathing. Call a POISON CENTER or doctor/physician if you feel unwell. Immediately call a POISON CENTER or doctor/physician. **IF SWALLOWED:** Call a POISON CENTER or doctor/physician if you feel unwell. Rinse mouth. Do NOT induce vomiting.

IN CASE OF FIRE: Use CO₂, dry chemical, or foam for extinction.

www.cancerdiagnostics.com

Cancer Diagnostics, Inc. 4300 Emperor Blvd., 400 Durham, NC 27703 (877) 846 5393 Fax: (877) 817 1716

59°F (15°C)  86°F (30°C)

REF CFQ100 1000mL

Store locked up at room temperature
Dispose of contents/container to an approved waste disposal plant.



MX1397 CalFor Fixative Liter AW-26034 Rev.C 2020-07-16

LOT



MADE IN USA

**COMPONENTS****CAS#**

Deionized Water 7732-18-5

Formic Acid 64-18-8

Formaldehyde 50-00-0

EMERGENCY CONTACT: Chemtrec 1-800-424-9300*For In Vitro Diagnostic Use Only*

Fixative Decalcifier **Fix** CalFor[®]

DIRECTIONS FOR USE:

Calfor is formulated to fix and decalcify simultaneously. Fresh tissue can be placed directly into Calfor; Needle biopsies may be completely fixed and decalcified in one hour or less. Small bone specimens and calcified arteries may be completely fixed and decalcified in 2-4 hours. An excellent decalcifier for large dense specimens for fixation and to begin decalcification. Calfor is a proven formulation for ISH and IHC.

DANGER: Harmful if swallowed. Harmful if inhaled. Causes severe skin burns and eye damage. May cause an allergic skin reaction. Obtain special instructions before use. Do not handle until all safety precautions have been read and understood. Use personal protective equipment as required. Wash face, hands and any exposed skin thoroughly after handling. Do not eat, drink or smoke when using this product. Use only outdoors or in a well-ventilated area. Do not breathe dust/fume/gas/mist/vapors/spray. Contaminated work clothing should not be allowed out of the workplace. Wear protective gloves.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or doctor/physician. **IF ON SKIN (OR HAIR):** Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. Wash contaminated clothing before reuse. If skin irritation or rash occurs: Get medical advice/attention. **IF INHALED:** Remove victim to fresh air and keep at rest in a position comfortable for breathing. Call a POISON CENTER or doctor/physician if you feel unwell. Immediately call a POISON CENTER or doctor/physician. **IF SWALLOWED:** Call a POISON CENTER or doctor/physician if you feel unwell. Rinse mouth. Do NOT induce vomiting.

IN CASE OF FIRE: Use CO₂, dry chemical, or foam for extinction.

www.cancerdiagnostics.com

Cancer Diagnostics, Inc. 4300 Emperor Blvd., 400 Durham, NC 27703 (877) 846 5393 Fax: (877) 817 1716



REF

CFQ400

1000mL CS/4

Store locked up at room temperature
Dispose of contents/container to an approved waste disposal plant.

MX1398 CalFor Fixative Liters, 4 AW-26034 Rev.C 2020-07-16

LOT



MADE IN USA