# **Aqua Mounter**



### **Intended Use**

For In Vitro Diagnostic Use.

## **Summary and Explanation**

Aqua Mounter is an aqueous-based mounting media designed for the preservation of Immunocytochemical and Immunohistochemical stainings.

Aqua Mounter is ideally suited for mounting cells or tissue sections immunostained with AEC, New Fuchsin, Fast Red, Alk Magenta, Alk Red and any Horseradish Peroxidase or Alkaline Phosphtase substrate-chromogens that are soluble in organic solvents.

#### **Presentation**

Aqua Mounter is provided in liquid form ready-to-use. It contains a buffer solution, reagents and a sodium azide anti-microbial

Catalog No.	Concentration	Volume/Qty		
BSB 0090	Ready-to-use	15 mL		
BSB 0091	Ready-to-use	50 mL		
BSB 0092	Ready-to-use	100 mL		
BSB 0093	Ready-to-use	500 mL		

Storage Store at 20-25°C

### **Stability**

**This product is stable up to the expiration date on the product label.** Do not use after expiration date listed on package label. Temperature fluctuations should be avoided. Store appropriately when not in use. Adhere to all local laws when disposing of this product.

#### **Precautions**

- 1 For professional users only. Results should be interpreted by a medical professional.
- $2. This product contains < 0.1\% so dium azide (NaN_3) as a preservative. Ensure proper handling procedures are used with this reagent.$
- 3. Always wear personal protective equipment such as laboratory coat, goggles and gloves when handling reagents.
- ${\bf 4.\, Dispose\, of\, unused\, solution\, according\, to\, local\, and\, federal\, regulations.}$
- 5. Do not ingest reagent. If reagent is ingested, seek medical advice immediately.
- 6. Avoid contact with eyes. If contact occurs, flush with large quantities of water.
- 7. For complete recommendations for handling biological specimens please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (1).

### **Preparation of Working Solution**

Aqua Mounter is a ready-to-use working solution and requires no further preparation.

PI 0093, Rev. E DCN: 3113

#### **Recommended Protocol**

- 1. After counterstaining and rinsing slides with buffer, apply 1-3 drops of Aqua Mounter to slide and coverslip.
- 2. Allow slide to dry at room temperature.
- 3. Observe under microscope.

### **Abbreviated Immunohistochemical Protocol**

Step ImmunoDetector AP/H		PolyDetector AP/HRP	PolyDetector Plus HRP		
Peroxidase/AP Blocker	5 min.	5 min.	5 min		
Primary Antibody	30-60 min.	30-60 min.	30-60 min.		
1st Step Detection	10 min.	30-45 min.	15 min. 15 min.		
2nd Step Detection	10 min.	Not Applicable			
Substrate-Chromogen	5-10 min.	5-10 min.	5-10 min.		
Counterstain / Coverslip	Varies	Varies	Varies		

### **Product Limitations**

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a qualified medical professional.

#### References

1. U.S. Department of Health and Human Services: Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. Supplement / Vol. 61, January 6, 2012.

## Symbol Key / Légende des symboles/Erläuterung der Symbole

_	Symbol Rey / Legende des symboles/Enddeerding der Symbole									
	EC RE	EMERGO EUROPE Prinsessegracht 20 2514 AP The Hague The Netherlands	arc arc	Storage Temperature Limites de température Zulässiger Temperaturbereich		Manufacturer Fabricant Hersteller	REF	Catalog Number Référence du catalogue Bestellnummer		
	IVD	In Vitro Diagnostic Medical Device Dispositif médical de diagnostic in vitro In-Vitro-Diagnostikum	$\bigcap_{\mathbf{i}}$	Read Instructions for Use Consulter les instructions d'utilisation Gebrauchsanweisung beachten		Expiration Date Utiliser jusque Verwendbar bis	LOT	Lot Number Code du lot Chargenbezeichnung		

