



Products for Anatomic Pathology

Product: 10% Neutral Buffered Formalin


Catalog Number: FX1000

Lot Number: 2334501

Date of Manufacture: 2023-12-15

Expiration Date: 2025-12-14

<u>TEST</u>	<u>SPECIFICATION</u>	<u>RESULT</u>
Appearance	Clear, colorless liquid	Passes test
Color (APHA)	<10	0
Assay.....	3.70% – 4.00%.....	3.80%
pH @ 25C	6.90 – 7.10	7.059



Jonathan D. Kent, Ph.D.
Director of Manufacturing