

09/15/2021

**Product Name:** 5 Gallon Pre-Fill Formalin

**Description:** 10% Neutral Buffered Formalin, 5 Gallon Cube, Cancer Diagnostics Brand,  
Each

**Product No.:** FX1000

**Lot No.:** 26218

**Date of Manufacture:** 2021-09-10

**Expiration Date:** 2023-09-01

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
<b>Container:</b>	CDC6378	03/31/221, 06/22/2021, 08/03/2021	N/A
<b>Spigot :</b>	COM4401	1348133	N/A
<b>Chemical:</b>	NBF-275G	1I001	2023-09-01
		1I002	2023-09-01
		1I003	2023-09-01


**QC Testing:**

LOT NUMBER: 1I001			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	9.85% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 1I002			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	9.70% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

LOT NUMBER: 11003			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.96 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex



**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**