



Products for Anatomic Pathology

Certificate of Analysis

FX1004 - 10% NBF, 7mL x 500/cs (7mL bottle w/5mL fill) (MTO)

Lot Number: 144327

Expiration Date: 2025-February-28

COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
pH @ 21 C - 28 C	≥ 6.8 and ≤ 7.2	6.8
Specific Gravity (g/cm ³)	≥ 1.014 and ≤ 1.019	1.015
Formalin Assay (%)	≥ 9 and ≤ 11	10.8

A handwritten signature in blue ink, appearing to read "Jonathan D. Kent", is written over a horizontal line.

Jonathan D. Kent, Ph.D.
Director of Manufacturing