
Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand

32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 23113

Date of Manufacture: 2020-12-18

Expiration Date: 2022-12-01

Materials:

 Part No.
 Lot No.
 Exp. Date

 ES147002
 K111120, L041120
 N/A

Container: ES147002 K111120, L041120 N/A **Lid**: ES147202 L060420, L070420, K160420 N/A

Chemical: NBF-275F 0L002 2022-12-01

QC Testing:

| LOT NUMBER: 0L002 | | | |
|----------------------------|----------------------------------|--------------|-------------|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| рН | 6.90-7.10 @ 25°C | 6.96 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.31% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com