

05/13/2021

Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand
32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 24696

Date of Manufacture: 2021-05-12

Expiration Date: 2023-04-01

Materials:


	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container:	ES147002	D191121, D081121	N/A
Lid:	ES147202	D090421	N/A
Chemical:	NBF-275F	1D011	2023-04-01

QC Testing:

LOT NUMBER: 1D011			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.99 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex



Jonathan D. Kent, Ph.D.
Director of Manufacturing