Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand

32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 24908

Date of Manufacture: 2021-05-27

Expiration Date: 2023-05-01

Materials:

 Part No.
 Lot No.
 Exp. Date

 ES147002
 D291121. B051121. B061121
 N/A

Container: ES147002 D291121, B051121, B061121 N/A **Lid**: ES147202 E010421, E020421 N/A

Chemical: NBF-275F 1E008 2023-05-01

QC Testing:

LOT NUMBER: 1E008			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.08 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com