Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand

32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 25155

Date of Manufacture: 2021-06-15

Expiration Date: 2023-06-01

Materials:

 Container:
 Part No. ES147002
 Lot No. D141121, D231121
 Exp. Date N/A

Lid: ES147202 D250421 N/A

Chemical: NBF-275F 1F004 2023-06-01

1F006 2023-06-01

QC Testina:

QC resting:			
LOT NUMBER: 1F004			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	7.09 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 1F006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.04 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com