Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand

32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 27355

Date of Manufacture: 2022-01-07

Expiration Date: 2023-12-01

Materials:

<u>Part No.</u> <u>Lot No.</u> <u>Exp. Date</u> iner: ES147002 J240821, J250821 N/A

Container: ES147002 J240821, J250821 N/A **Lid**: ES147202 J010421, J020421, J030421 N/A

Chemical: NBF-275F 1L012 2023-12-01

QC Testina:

LOT NUMBER: 1L012			
PROPERTY	SPECIFICATION	RESULTS	PASS/FAIL
pH	6.90-7.10 @ 25°C	7.04 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	9.85% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com