Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand

32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 27541

Date of Manufacture: 2022-01-27

Expiration Date: 2024-01-01

Materials:

 Container:
 Part No. ES147002
 Lot No. A081222
 Exp. Date N/A

Lid: ES147202 K070421, K080421 N/A

Chemical: NBF-275F 2A005 2024-01-01

QC Testing:

LOT NUMBER: 2A005			
PROPERTY	SPECIFICATION	RESULTS	PASS/FAIL
pH	6.90-7.10 @ 25°C	6.99 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com