Certificate of Analysis

04/06/2022



Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand 32 vials per pack, 24 packs/case

Product No.: FX1006-HPY-768

Lot No.: 28514

Date of Manufacture: 2022-04-05

Expiration Date: 2024-03-01

Materials:

| Container: | <u>Part No.</u> | <u>Lot No.</u> | <u>Exp. Date</u> |
|------------|-----------------|---------------------------|------------------|
| | ES147002 | A231222, A241222, C230222 | N/A |
| | ES147202 | B050422, B060422, B070422 | N/A |
| | NBF-275F | 2C017 | 2024-03-01 |

QC Testing:

| LOT NUMBER: 2C017 | | | | | |
|----------------------------|-------------------------------------|--------------|-------------|--|--|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL | | |
| рН | 6.90-7.10 @ 25°C | 7.02 @ 25°C | PASS | | |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.46% (v/v) | PASS | | |
| Appearance | Clear with No Particulate | Clear | PASS | | |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS | | |

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing