Certificate of Analysis



06/01/2021

Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand

32 vials per pack, 4 packs/case

Product No.: FX1006-HPY

Lot No.: 24909

Date of Manufacture: 2021-05-28

Expiration Date: 2023-05-01

Materials:

 Container
 ES147002
 Lot No. B051121, B061121, B131121
 Exp. Date N/A

 Lid:
 ES147202
 E010421, E050421, Q040421
 N/A

Chemical: NBF-275F 1E008 2023-05-01

1E009 2023-05-01

QC Testina:

| QC resuing. | | | |
|----------------------------|-----------------------------------|--------------|-------------|
| LOT NUMBER: 1E008 | | | |
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| рН | 6.90-7.10 @ 25°C | 7.08 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.16% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |
| LOT NUMBER: 1E009 | | | |
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| рН | 6.90-7.10 @ 25°C | 7.05 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 9.85% (v/v) | PASS |
| Appearance \ \ | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Jonathan D. Kent, Ph.D.

Country of Original United States of America Latex Free: This product contains no latex

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com