Certificate of Analysis

02/28/2022



Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand 32 vials per pack, 4 packs/case

Product No.: FX1006-HPY

Lot No.: 27931

Date of Manufacture: 2022-02-25

Expiration Date: 2024-02-01

Materials:

	Part No.	Lot No.	Exp. Date
Container	ES147002	I220821, J030821, L080821, L090821	N/A
Lid:	ES147202	L100421, I270421, I280421, I290421	N/A
Chemical:	NBF-275F	2B010	2024-02-01

QC Testing:

LOT NUMBER: 2B010				
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL	
рН	6.90-7.10 @ 25°C	6.93 @ 25°C	PASS	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS	
Appearance	Clear with No Particulate	Clear	PASS	
Labeling	Correct Lot Number and Exp. Date	Correct	PASS	

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing