Certificate of Analysis

09/09/2022



Product Name: 20mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 20 ml half filled, Cancer Diagnostics Brand 32 vials per pack, 4 packs/case

Product No.: FX1006-HPY

Lot No.: 30327

Date of Manufacture: 2022-09-08

Expiration Date: 2024-08-01

Materials:

	Part No.	Lot No.	Exp. Date
Container	ES147002	G180222	N/A
Lid:	ES147202	G290422, G300422, G310422	N/A
Chemical:	NBF-275F	2H010	2024-08-01
		21001	2024-09-01

QC Testing:

LOT NUMBER: 2H010					
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL		
рН	6.90-7.10 @ 25°C	6.99 @ 25°C	PASS		
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS		
Appearance	Clear with No Particulate	Clear	PASS		
Labeling	Correct Lot Number and Exp. Date	Correct	PASS		
LOT NUMBER: 21001					
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL		
рН	6.90-7.10 @ 25°C	6.95 @ 25°C	PASS		
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS		
Appearance	Clear with No Particulate	Clear	PASS		
Labeling	Correct Lot Number and Exp. Date	Correct	PASS		

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing