



Certificate of Analysis

2/3/2023

Product No.: FX1006-HPY

Description: Cancer Diagnostics labeled 10% Neutral Buffered Formalin, Prefilled Container, 20ml, 32 vials per pack 128 per case

Lot No.: 31934

Date of Manufacturing: 2/2/2023


Expiration Date: 1/1/2025

10% Formalin Lot #: 3A011

PROPERTY	TEST SPECIFICATION	RESULTS
PH	6.90-7.10 @ 25°C	7.1@ 25°C
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31%(v/v)
Appearance	Clear, colorless, no observable particulate	Conforms

Country of Origin: United States of America

Latex Free: This product contains no latex



Jonathan D. Kent, Ph.D.
Director of Manufacturing