## **Certificate of Analysis**



7/14/2023

Product No.: FX1006-HPY
Description: Cancer Diagnostics labeled 10% Neutral Buffered Formalin, Prefilled Container, 20ml, 32 vials per pack 128 per case
Lot No.: 34066
Date of Manufacturing: 7/14/2023
Expiration Date: 7/1/2025

## **10% Formalin Lot #: 3G002**

Specification Appearance Formaldehyde Concentration pH @ 25°C **Requirements** Clear, colorless, no observable particulate 9.5-10.5%(v/v) 6.90-7.10 **Results** Conforms 10.00% (v/v) 7.08

**Country of Origin:** United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing