Certificate of Analysis



11/7/2023

Product No.: FX1006-HPY

Description: Cancer Diagnostics labeled 10% Neutral Buffered Formalin, Prefilled Container, 20ml, 32 vials per pack

128 per case

Lot No.: 35528

Date of Manufacturing: 11/1/2023

Expiration Date: 10/1/2025

| 10% Formalin Lot #: 3J012 | |
|---|-------------------------|
| Requirements | Results |
| Clear, colorless, no observable particulate | Clear liquid - Conforms |

Formaldehyde Concentration pH @ 25°C

SpecificationAppearance

Clear, colorless, no observable particulate 9.50 -10.50%(v/v) 6.90 - 7.10

10.16%(v/v) 6.90

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com