Certificate of Analysis

3/7/2024



Products for Anatomic Pathology

Product No.: FX1006-HPY

Description: Cancer Diagnostics labeled 10% Neutral Buffered Formalin, Prefilled Container, 20ml, 32 vials per pack

128 per case

Lot No.: 37160

Date of Manufacturing: 3/5/2024

Expiration Date: 2/1/2026

10% Formalin Lot #: 4B022			
Specification	Requirements	Results	
Appearance	Clear, colorless, no observable particulate	Clear liquid - Conforms	
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.00% (v/v)	
pH @ 25°C	6.90 - 7.10	7.08	

Country of Origin: United States of America **Latex Free:** This product contains no latex

> Jonathan D. Kent, Ph.D. **Director of Manufacturing**

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com