Certificate of Analysis

08/27/2021



Product Name: 40mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 40 ml half filled, Cancer Diagnostics Brand 24 vials per pack, 2 packs/case

Product No.: FX1007-HPY

Lot No.: 25877

Date of Manufacture: 2021-08-26

Expiration Date: 2023-07-01

Materials:

Container	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
	ES146002	A311221	N/A
	ES144202 NBF-275F		N/A 2023-07-01

QC Testing:

LOT NUMBER: 1G006					
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL		
рН	6.90-7.10 @ 25°C	7.03 @ 25°C	PASS		
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS		
Appearance	Clear with No Particulate	Clear	PASS		
Labeling	Correct Lot Number and Exp. Date	Correct	PASS		

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing