



Products for Anatomic Pathology

# Certificate of Analysis

09/23/2021

**Product Name:** 40mL Pre-Fill Formalin

**Description:** 10% Neutral Buffered Formalin, Prefilled Container, 40 ml half filled, Cancer Diagnostics Brand  
24 vials per pack, 2 packs/case

**Product No.:** FX1007-HPY

**Lot No.:** 26330

**Date of Manufacture:** 2021-09-22

**Expiration Date:** 2023-09-01

**Materials:**

|                  | <u>Part No.</u> | <u>Lot No.</u>            | <u>Exp. Date</u> |
|------------------|-----------------|---------------------------|------------------|
| <b>Container</b> | ES146002        | A311221, B011121, H111221 | N/A              |
| <b>Lid:</b>      | ES144202        | D201221, D211221          | N/A              |
| <b>Chemical:</b> | NBF-275F        | 1I002                     | 2023-09-01       |

**QC Testing:**

| <b>LOT NUMBER: 1I002</b>          |                                  |                |                    |
|-----------------------------------|----------------------------------|----------------|--------------------|
| <b>PROPERTY</b>                   | <b>SPECIFICATION</b>             | <b>RESULTS</b> | <b>PASS / FAIL</b> |
| <b>pH</b>                         | 6.90-7.10 @ 25°C                 | 6.98 @ 25°C    | <b>PASS</b>        |
| <b>Formaldehyde Concentration</b> | 9.5-10.5% (v/v)                  | 9.70% (v/v)    | <b>PASS</b>        |
| <b>Appearance</b>                 | Clear with No Particulate        | Clear          | <b>PASS</b>        |
| <b>Labeling</b>                   | Correct Lot Number and Exp. Date | Correct        | <b>PASS</b>        |

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**