

08/18/2022

Product Name: 40mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 40 ml half filled, Cancer Diagnostics Brand
24 vials per pack, 2 packs/case

Product No.: FX1007-HPY

Lot No.: 30028

Date of Manufacture: 2022-08-17

Expiration Date: 2024-08-01

Materials:


| | <u>Part No.</u> | <u>Lot No.</u> | <u>Exp. Date</u> |
|------------------|-----------------|---------------------------|------------------|
| Container | ES146002 | H121221, H131221 | N/A |
| Lid: | ES144202 | F261222, F271222, E221222 | N/A |
| Chemical: | NBF-275F | 2H002 | 2024-08-01 |

QC Testing:

| LOT NUMBER: 2H002 | | | |
|----------------------------|----------------------------------|--------------|-------------|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| pH | 6.90-7.10 @ 25°C | 7.06 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.46% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Country of Origin: United States of America

Latex Free: This product contains no latex



Jonathan D. Kent, Ph.D.
Director of Manufacturing