Certificate of Analysis





Product Name: 60mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 60 ml half filled, Cancer Diagnostics Brand

27 vials per pack, 2 packs/case

Product No.: FX1008-HPY

Lot No.: 23234

Date of Manufacture: 2020-12-31

Expiration Date: 2022-12-01

Materials:

 Part No.
 Lot No.
 Exp. Date

 Container
 ES145003
 H170820, K140820
 N/A

 Lid:
 ES144202
 J271220, J281220
 N/A

 Chemical:
 NPF 3755
 01 000
 3033 13 01

Chemical: NBF-275F 0L009 2022-12-01

QC Testing:

LOT NUMBER: 0L009			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	6.91 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com