
Certificate of Analysis





Product Name: 60mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 60 ml half filled, Cancer Diagnostics Brand

27 vials per pack, 2 packs/case

Product No.: FX1008-HPY

Lot No.: 25412

Date of Manufacture: 2021-07-07

Expiration Date: 2023-06-01

Materials:

 Part No.
 Lot No.
 Exp. Date

 Container
 ES145003
 E160221
 N/A

 Lid:
 ES144202
 D191221
 N/A

 Chemical:
 NBF-275F
 1F013
 2023-06-01

QC Testing:

LOT NUMBER: 1F013			
PROPERTY	SPECIFICATION	RESULTS	PASS/FAIL
рН	6.90-7.10 @ 25°C	7.07 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

> Jonathan D. Kent, Ph.D. Director of Manufacturing