## **Certificate of Analysis**

07/02/2021



Product Name: 90mL Pre-Fill Formalin

**Description**: 10% Neutral Buffered Formalin, Prefilled Container, 90 ml half filled, Cancer Diagnostics Brand 12 vials per pack, 4 packs/case

Product No.: FX1009-HPY

Lot No.: 25253

Date of Manufacture: 2021-07-01

Expiration Date: 2023-06-01

Materials:

	Part No.	Lot No.	Exp. Date
Container	ES144002	D080221	N/A
Lid:	ES144202	D211221, F181221	N/A
Chemical:	NBF-275F	1F011	2023-06-01

## **QC** Testing:

LOT NUMBER: 1F011					
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL		
рН	6.90-7.10 @ 25°C	6.95 @ 25°C	PASS		
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS		
Appearance	Clear with No Particulate	Clear	PASS		
Labeling	Correct Lot Number and Exp. Date	Correct	PASS		

**Country of Origin:** United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing