Certificate of Analysis





Product Name: 90mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 90 ml half filled, Cancer Diagnostics Brand

12 vials per pack, 4 packs/case

Product No.: FX1009-HPY

Lot No.: 25686

Date of Manufacture: 2021-07-28

Expiration Date: 2023-07-01

Materials:

 Container
 Est No. Est 144002
 Lot No. D130221
 Exp. Date N/A

 Lid:
 Est 144202
 F221221
 N/A

 Chemical:
 NBF-275F
 1G006
 2023-07-01

QC Testing:

LOT NUMBER: 1G006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.03 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com