



Products for Anatomic Pathology

Certificate of Analysis

FX1010-70 - 10% NBF, 120mL, (96 p/CS) 70mL fill line (MTO)

Lot Number: 149678

Expiration Date: 2025-May-31

COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
pH @ 21 C - 28 C	≥ 6.8 and ≤ 7.2	6.9
Specific Gravity (g/cm ³)	≥ 1.014 and ≤ 1.019	1.015
Formalin Assay (%)	≥ 9.0 and ≤ 11.0	10.5

Jonathan D. Kent, Ph.D.
Director of Manufacturing