Certificate of Analysis



06/15/2021

Product Name: 120mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 120 ml half filled, Cancer Diagnostics Brand

24 vials per pack, 2 packs/case

Product No.: FX1010-HPY

Lot No.: 24928

Date of Manufacture: 2021-06-14

Expiration Date: 2023-06-01

Materials:

 Container
 Part No. ES142002
 Lot No. D100821
 Exp. Date N/A

 Lid:
 ES143202
 D170521, P160521
 N/A

Chemical: NBF-275F 1F005 2023-06-01

QC Testing:

LOT NUMBER: 1F004			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	7.03 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com