Certificate of Analysis



Products for Anatomic Pathology

09/28/2021

Product Name: 120mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 120 ml half filled, Cancer Diagnostics Brand

24 vials per pack, 2 packs/case

Product No.: FX1010-HPY

Lot No.: 26331

Date of Manufacture: 2021-09-27

Expiration Date: 2023-09-01

Container

Materials:

Part No. Lot No. ES142002 E270821

Lid: ES143202 S251221 N/A

Chemical: NBF-275F 11002 2023-09-01

QC Testing:

LOT NUMBER: 11002			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	9.70% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

> Jonathan D. Kent, Ph.D. Director of Manufacturing

Exp. Date

N/A

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com