Certificate of Analysis

10/18/2022



Product Name: 120mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 120 ml half filled, Cancer Diagnostics Brand 24 vials per pack, 2 packs/case

Product No.: FX1010-HPY

Lot No.: 30747

Date of Manufacture: 2022-10-18

Expiration Date: 2024-10-01

Materials:

Container	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
	FS142002	E200822, K140221	N/A
Lid:		E220522, B270222	N/A 2024-10-01

QC Testing:

LOT NUMBER: 2J003				
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL	
рН	6.90-7.10 @ 25°C	7.03 @ 25°C	PASS	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS	
Appearance	Clear with No Particulate	Clear	PASS	
Labeling	Correct Lot Number and Exp. Date	Correct	PASS	

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing