



Products for Anatomic Pathology

# Certificate of Analysis

05/26/2021

**Product Name:** 240mL Pre-Fill Formalin

**Description:** 10% Neutral Buffered Formalin, Prefilled Container, 240 ml half filled, Cancer Diagnostics Brand 24 jars/case

**Product No.:** FX1011-24

**Lot No.:** 24914

**Date of Manufacture:** 2021-05-24

**Expiration Date:** 2023-05-01

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
<b>Container</b>	08-070-SWPP	202103601572	N/A
<b>Lid:</b>	070-219-115	202102599237	N/A
<b>Chemical:</b>	NBF-275F	1E005	2023-05-01

**QC Testing:**

**LOT NUMBER: 1E005**

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.01 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**