Certificate of Analysis



02/18/2022

Product Name: 240mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 240 ml half filled, Cancer Diagnostics Brand

24 jars/case

Product No.: FX1011-24

Lot No.: 27971

Date of Manufacture: 2022-02-17

Expiration Date: 2024-02-01

Materials:

Part No. Lot No. Exp. Date

Container 08-070- 202112624840 N/A

SWPP

Lid: 070-219- L400HQW N/A

115-PC

Chemical: NBF-275F 2B006 2024-02-01

QC Testing:

LOT NUMBER: 2B006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	7.07 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com