


FX1020-96 - 10% Zinc Formalin, Cancer Diagnostics, 20mL, 96/cs (MTO)

Lot Number: 132145

Expiration Date: 2023-October-31

COMPONENTS

Test	Specifications	Results
Color	Clear	Clear
pH @ 22 - 25 C	≥ 5.6 and ≤ 5.8	5.8
Specific Gravity (g/cm ³)	≥ 0.980 and ≤ 1.030	1.010
Formalin Assay (%)	≥ 9.0 and ≤ 11.0	9.3



Jonathan D. Kent, Ph.D.
Director of Manufacturing