



**Product Name:** 480 mL Pre-Fill Formalin

**Description:** 10% Neutral Buffered Formalin, Prefilled Container, 480 ml half filled, Cancer Diagnostics Brand 24/case

**Product No.:** FX1116-24

**Lot No.:** 27188

**Date of Manufacture:** 2021-12-15

**Expiration Date:** 2023-12-01

**Materials:**


|                   | <u>Part No.</u> | <u>Lot No.</u> | <u>Exp. Date</u> |
|-------------------|-----------------|----------------|------------------|
| <b>Container:</b> | 16-089-SWPP     | 202110620165   | N/A              |
| <b>Lid:</b>       | 089-828-115     | 202110621071   | N/A              |
| <b>Chemical:</b>  | NBF-275F        | 1L003          | 2023-12-01       |

**QC Testing:**

| <b>LOT NUMBER: 1L003</b>   |                                  |                |                    |
|----------------------------|----------------------------------|----------------|--------------------|
| <b>PROPERTY</b>            | <b>SPECIFICATION</b>             | <b>RESULTS</b> | <b>PASS / FAIL</b> |
| pH                         | 6.90-7.10 @ 25°C                 | 6.96 @ 25°C    | <b>PASS</b>        |
| Formaldehyde Concentration | 9.5-10.5% (v/v)                  | 10.31% (v/v)   | <b>PASS</b>        |
| Appearance                 | Clear with No Particulate        | Clear          | <b>PASS</b>        |
| Labeling                   | Correct Lot Number and Exp. Date | Correct        | <b>PASS</b>        |

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex



**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**