

09/07/2022

Product Name: 480 mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 480 ml half filled, Cancer Diagnostics Brand
24/case

Product No.: FX1116-24

Lot No.: 30281

Date of Manufacture: 2022-09-06

Expiration Date: 2024-08-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container:	16-089-SWPP	202206639089	N/A
Lid:	089-828-115	202205636685	N/A
Chemical:	NBF-275F	2H010	2024-08-01

QC Testing:

LOT NUMBER: 2H010			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.99 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.16% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex



Jonathan D. Kent, Ph.D.
Director of Manufacturing