



Products for Anatomic Pathology

Certificate of Analysis

01/29/2021

Product Name: 1000mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 1000 ml half filled, Cancer Diagnostics Brand 12/case

Product No.: FX1132-12

Lot No.: 23522

Date of Manufacture: 2021-01-29

Expiration Date: 2023-01-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container:	32-120-SWPP	202012594408	N/A
Lid:	120-828-115	202012593432	N/A
Chemical:	NBF-275F	1A006	2023-01-01
		1A007	2023-01-01

QC Testing:

LOT NUMBER: 1A006

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.00 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

LOT NUMBER: 1A007

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America
Latex Free: This product contains no latex

Jonathan D. Kent, Ph.D.
Director of Manufacturing