Certificate of Analysis



06/13/2022



Product Name: 1000mL Pre-Fill Formalin

Description: 10% Neutral Buffered Formalin, Prefilled Container, 1000 ml half filled, Cancer Diagnostics Brand

12/case

Product No.: FX1132-12

Lot No.: 29397

Date of Manufacture: 2022-06-13

Expiration Date: 2024-05-01

Materials:

Part No. Lot No. Exp. Date Container: 32-120-SWPP 202202630156 N/A

Lid: 120-828-115 202112625449 N/A

Chemical: NBF-275F 2E010 2024-03-01

QC Testina:

LOT NUMBER: 2E010			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	7.09 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

> Jonathan D. Kent, Ph.D. **Director of Manufacturing**

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com