

06/18/2021

**Product Name:** 20mL Pre-Fill Formalin

**Description:** Kit, 8 place Prostate Biopsy

**Product No.:** PP1008

**Lot No.:** 25221

**Date of Manufacture:** 2021-06-17

**Expiration Date:** 2023-06-01

**Materials:**


	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
<b>Container</b>	ES147002	D131121, D151121, D161121	N/A
<b>Lid:</b>	ES147202	E100421, E110421	N/A
<b>Chemical:</b>	NBF-275F	1F006	2023-06-01

**QC Testing:**

LOT NUMBER: 1F006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.04 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex



**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**