

**Product Name:** 20mL Pre-Fill Formalin

**Description:** Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

**Product No.:** PP1012

**Lot No.:** 26281

**Date of Manufacture:** 2021-09-20

**Expiration Date:** 2023-09-01

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container	ES147002	G181121, G191121	N/A
Lid:	ES147202	F250221	N/A
Chemical:	NBF-275F	11004	2023-09-01


**QC Testing:**

**LOT NUMBER: 11004**

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.02 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex



**Jonathan D. Kent, Ph.D.**  
Director of Manufacturing