

02/15/2022

Product Name: 20mL Pre-Fill Formalin

Description: Rtquvcg'Rcem 12 place Prostate Biopsy'Mkv
'''

Product No.: PP1012

Lot No.: 27782

Date of Manufacture: 2022-02-10

Expiration Date: 2024-02-01

Materials:


	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container	ES147002	A181222, A191222	N/A
Lid:	ES147202	J170421, J180421	N/A
Chemical:	NBF-275F	2B001	2024-02-01

QC Testing:

LOT NUMBER: 2B001			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.01 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex


Jonathan D. Kent, Ph.D.
Director of Manufacturing