Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10%

formalin per kit, 25 kits per case

Product No.: PP1012

Lot No.: 28241

Date of Manufacture: 2022-03-23

Expiration Date: 2024-03-01

Materials:

 Container
 Part No. ES147002
 Lot No. A131222
 Exp. Date N/A

 Lid:
 ES147202
 A260422
 N/A

 Chemical:
 NBF-275F
 2C011
 2024-03-01

QC Testing:

| LOT NUMBER: 2C011 | | | |
|----------------------------|----------------------------------|--------------|-------------|
| PROPERTY | SPECIFICATION | RESULTS | PASS / FAIL |
| pH | 6.90-7.10 @ 25°C | 7.02 @ 25°C | PASS |
| Formaldehyde Concentration | 9.5-10.5% (v/v) | 10.46% (v/v) | PASS |
| Appearance | Clear with No Particulate | Clear | PASS |
| Labeling | Correct Lot Number and Exp. Date | Correct | PASS |

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com