

04/21/2022

**Product Name:** 20mL Pre-Fill Formalin

**Description:** Kit, 12 place Prostate Biopsy

**Product No.:** PP1012

**Lot No.:** 28823

**Date of Manufacture:** 2022-04-20

**Expiration Date:** 2024-04-01

**Materials:**


	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container	ES147002	D010222	N/A
Lid:	ES147202	C250422	N/A
Chemical:	NBF-275F	2D005	2024-04-01

**QC Testing:**

LOT NUMBER: 2D005			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.92 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex



**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**