

07/07/2022

**Product Name:** 20mL Pre-Fill Formalin

**Description:** Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

**Product No.:** PP1012

**Lot No.:** 29690

**Date of Manufacture:** 2022-07-07

**Expiration Date:** 2024-06-01

**Materials:**


	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container	ES147002	D270222	N/A
Lid:	ES147202	E210422, E220422	N/A
Chemical:	NBF-275F	2F007	2024-06-01

**QC Testing:**

LOT NUMBER: 2F007			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.10 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex



**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**