
Certificate of Analysis

08/18/2022



Product Name: 20mL Pre-Fill Formalin

Products for Anatomic Pathology

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10%

formalin per kit, 25 kits per case

Product No.: PP1012

Lot No.: 30112

Date of Manufacture: 2022-08-18

Expiration Date: 2024-08-01

Materials:

 Container
 Part No. ES147002
 Lot No. G100222, G120222
 Exp. Date N/A

 Lid:
 ES147202
 F290422
 N/A

 Chemical:
 NBF-275F
 2H004
 2024-08-01

QC Testing:

LOT NUMBER: 2H004			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	7.01 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com