

11/07/2022

Product Name: 20mL Pre-Fill Formalin

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

Product No.: PP1012

Lot No.: 31010

Date of Manufacture: 2022-11-03

Expiration Date: 2024-10-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container	ES147002	I201122	N/A
Lid:	ES147202	H250422	N/A
Chemical:	NBF-275F	2J009	2024-10-01

QC Testing:

LOT NUMBER: 2J009			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex



Jonathan D. Kent, Ph.D.
Director of Manufacturing