Certificate of Analysis





Product Name: 20mL Pre-Fill Formalin

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10%

formalin per kit, 25 kits per case

Product No.: PP1012

Lot No.: 31197

Date of Manufacture: 2022-11-28

Expiration Date: 2024-11-01

Materials:

Part No. Lot No. Exp. Date Container ES147002 1271122, 1281122 N/A Lid: ES147202 H240422 N/A Chemical: NBF-275F 2K006 2024-11-01

QC Testing:

LOT NUMBER: 2K006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
рН	6.90-7.10 @ 25°C	6.97 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America Latex Free: This product contains no latex

> Jonathan D. Kent, Ph.D. **Director of Manufacturing**

Tel: (877) 846-5393 | Fax: (877) 817-1716 | www.cancerdiagnostics.com