Certificate of Analysis



12/30/2022

Product No.: PP1012

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10%

formalin per kit, 25 kits per case

Lot No.: 31376

Date of Manufacturing: 12/30/2022

Expiration Date: 12/1/2024

10% Formalin Lot #: 2L009		
PROPERTY	TEST SPECIFICATION	RESULTS
PH	6.90-7.10 @ 25°C	7.02@ 25°C
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31%(v/v)
Appearance	Clear, colorless, no observable particulate	Conforms

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing