Certificate of Analysis



5/17/2023

Product No.: PP1012Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per caseLot No.: 33266

Date of Manufacturing: 5/16/2023

Expiration Date: 5/1/2025

10% Formalin Lot #: 3E006

Specification Appearance Formaldehyde Concentration pH @ 25°C **Requirements** Clear, colorless, no observable particulate 9.5-10.5%(v/v) 6.90-7.10

Results Conforms 10.31%(v/v) 6.93

Country of Origin: United States of America **Latex Free:** This product contains no latex

Jonathan D. Kent, Ph.D. Director of Manufacturing