



7/31/2023

Product No.: PP1012

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

Lot No.: 33974

Date of Manufacturing: 7/31/2023

Expiration Date: 7/1/2025

10% Formalin Lot #: 3G011

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.5-10.5%(v/v)	9.85%(v/v)
pH @ 25°C	6.90-7.10	6.96

Country of Origin: United States of America

Latex Free: This product contains no latex

Signature of Jonathan D. Kent, Ph.D., Director of Manufacturing